

Chemical Analysis Report

NW-DIST-2022-04-19-02

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **Milton South Run**
Request ID: **RQ-2022-03-07-52**
Customer: **NW-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
160 Government Center
Suite 208
Pensacola, FL 32502-5794
Attn: Myranda Butler

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Colin Wright, Program Administrator

Date Certified: 09-MAY-2022 11:01

A handwritten signature in black ink, appearing to read "Colin Wright", is displayed on a light gray rectangular background.

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical groups are included in this report: Metals, Nutrients and Pesticides.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited by the Florida Department of Health Environmental Laboratory Certification Program or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Jakes Bayou upstream of Robinson Point Rd Collection Date/Time: 04/18/2022 10:25

Field ID: G4NW0394

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2320186	EPA 180.1 Rev. 2.0	Turbidity	5.4		NTU	P412573	
	SM 2320 B-2011	Total Alkalinity	5.5		mg CaCO3/L	P413141	
	SM 2540 D-2011	TSS	4	I	mg/L	P412991	
	SM 4500-F- C-2011	Fluoride	0.039	I	mg F/L	P413148	
2320187	EPA 350.1 Rev. 2.0	Ammonia-N	0.016		mg N/L	P412909	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.31		mg N/L	P413093	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.014		mg N/L	P412789	
	EPA 365.1 Rev. 2.0	Total-P	0.020		mg P/L	P412799	
	SM 5310 B-2011	Organic Carbon	7.1		mg C/L	P412742	
2320188	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P412555	

Ref. Method and Comment:

SM 2540 D-2011: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Sample Location: Sandy Point Creek below Sandy Point Street

Collection Date/Time: 04/18/2022 11:10

Field ID: G4NW0080

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2320189	DEP SOP: NU-094-1	Color (true)**	280		PCU	P412564	
	EPA 180.1 Rev. 2.0	Turbidity	5.4		NTU	P412573	
	EPA 300.0 Rev. 2.1	Chloride	12		mg Cl/L	P412954	
		Sulfate	1.1		mg SO4/L	P412957	
	SM 2320 B-2011	Total Alkalinity	8.4		mg CaCO3/L	P413141	
	SM 2540 C-2011	TDS	72		mg/L	P413063	
	SM 2540 D-2011	TSS	6	I	mg/L	P412990	
	SM 4500-F- C-2011	Fluoride	0.035	U	mg F/L	P413148	
2320190	EPA 350.1 Rev. 2.0	Ammonia-N	0.016		mg N/L	P412909	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.81		mg N/L	P413093	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.006	I	mg N/L	P412790	
	EPA 365.1 Rev. 2.0	Total-P	0.024		mg P/L	P412799	
	SM 5310 B-2011	Organic Carbon	19		mg C/L	P412742	
2320191	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P412555	
2320197	EPA 8321B	2,4-D	0.0020	U	ug/L	P412554	
		Acesulfame K	0.12	I	ug/L	P412593	
		2,4,5-T	0.0040	U	ug/L	P412554	
		AMPA	0.10	U	ug/L	P412593	
		Acetaminophen	0.0080	U	ug/L	P412554	
		Acetamiprid	0.0020	U	ug/L	P412554	
		Endothall	0.25	U	ug/L	P412593	
		Glufosinate	0.10	U	ug/L	P412593	
		Glyphosate	0.10	U	ug/L	P412593	
		Afidopyrophen	0.0020	U	ug/L	P412554	
		Bentazon	0.0038		ug/L	P412554	
		Sucralose	0.70		ug/L	P412593	
		Benzovindiflupyr	0.0020	U	ug/L	P412554	
		Carbamazepine	8.0E-04	U	ug/L	P412554	
		Clothianidin	0.0020	U	ug/L	P412554	
		Dinotefuran	0.0020	U	ug/L	P412554	
		Diuron	0.0020	U	ug/L	P412554	
		Fenuron	0.032	U	ug/L	P412554	
		Fluridone	8.0E-04	U	ug/L	P412554	
		Imazapyr	0.0056	I	ug/L	P412554	
		Hydrocodone	0.0020	U	ug/L	P412554	
		Ibuprofen	0.020	U	ug/L	P412554	
		Imidacloprid	0.0032	I	ug/L	P412554	MS
		Linuron	0.0040	U	ug/L	P412554	
		Mandestrobin	0.0020	U	ug/L	P412554	
		MCP	0.0021	I	ug/L	P412554	
		Naproxen	0.0080	U	ug/L	P412554	
		Primidone	0.0040	U	ug/L	P412554	
		Pyraclostrobin	0.0020	U	ug/L	P412554	

Field ID: G4NW0080

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2320197	EPA 8321B	Silvex	0.0040	U	ug/L	P412554	RPD
		Thiamethoxam	0.0020	U	ug/L	P412554	MS
		Tolfenpyrad	0.0020	U	ug/L	P412554	
		Triclopyr	0.0040	U	ug/L	P412554	
2320198	EPA 200.7 Rev. 4.4	Calcium	3.38		mg/L	P412588	
		Iron	1.60E+03		ug/L	P412588	
		Magnesium	0.98		mg/L	P412588	
		Potassium	1.8		mg/L	P412588	
		Sodium	8.9		mg/L	P412588	
		Strontium	9.2		ug/L	P412588	
		Tin	3.0	U	ug/L	P412588	
		Titanium	2.1	I	ug/L	P412588	
		Vanadium	2.0	U	ug/L	P412588	
		Zinc	5.0	U	ug/L	P412588	
	EPA 200.8 Rev. 5.4	Aluminum	386		ug/L	P412588	
		Antimony	0.050	U	ug/L	P412588	
		Arsenic	0.65		ug/L	P412588	
		Barium	6.04		ug/L	P412588	
		Beryllium	0.015	I	ug/L	P412588	
		Boron**	53.7		ug/L	P412588	
		Cadmium	0.020	U	ug/L	P412588	
		Chromium	0.50	I	ug/L	P412588	
		Cobalt	0.183		ug/L	P412588	
		Copper	0.40	U	ug/L	P412588	
		Lead	0.69		ug/L	P412588	
		Manganese	10.3		ug/L	P412588	
		Molybdenum	0.15	U	ug/L	P412588	
		Nickel	0.54	I	ug/L	P412588	
		Selenium	0.20	U	ug/L	P412588	
		Silver	0.010	U	ug/L	P412588	
		Thallium	0.10	U	ug/L	P412588	

Ref. Method and Comment:

SM 2540 D-2011: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

EPA 200.8 Rev. 5.4: The lead result was confirmed in the undigested sample on 04/22/2022.

Quality Assurance Report Method Blank Results

Reference Method: DEP SOP: NU-094-1
Batch ID: P412564

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P412573

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P412588

Component	Result	Code	Units
Calcium	0.075	U	mg/L
Iron	30	U	ug/L
Magnesium	0.040	U	mg/L
Potassium	0.30	U	mg/L
Sodium	0.50	U	mg/L
Strontium	2.0	U	ug/L
Tin	3.0	U	ug/L
Titanium	0.75	U	ug/L
Vanadium	2.0	U	ug/L
Zinc	5.0	U	ug/L

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P412588

Component	Result	Code	Units
Aluminum	5.0	U	ug/L
Antimony	0.050	U	ug/L
Arsenic	0.050	U	ug/L
Barium	0.20	U	ug/L
Beryllium	0.010	U	ug/L
Boron	2.0	U	ug/L
Cadmium	0.020	U	ug/L
Chromium	0.40	U	ug/L
Cobalt	0.020	U	ug/L
Copper	0.40	U	ug/L
Lead	0.20	U	ug/L
Manganese	0.10	U	ug/L
Molybdenum	0.15	U	ug/L
Nickel	0.50	U	ug/L
Selenium	0.20	U	ug/L
Silver	0.010	U	ug/L
Thallium	0.10	U	ug/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P412954

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Quality Assurance Report Method Blank Results

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P412957

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P412909

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P413093

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P412789

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P412790

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P412799

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P412555

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: EPA 8321B
Batch ID: P412554

Component	Result	Code	Units
2,4-D	0.0020	U	ug/L
2,4,5-T	0.0040	U	ug/L
Acetaminophen	0.089		ug/L
Acetamiprid	0.0020	U	ug/L
Afidopyropen	0.0020	U	ug/L
Bentazon	8.0E-04	U	ug/L
Benzovindiflupyr	0.0020	U	ug/L
Carbamazepine	8.0E-04	U	ug/L
Clothianidin	0.0020	U	ug/L
Dinotefuran	0.0020	U	ug/L

Quality Assurance Report Method Blank Results

Reference Method: EPA 8321B

Batch ID: P412554

Component	Result	Code	Units
Diuron	0.0020	U	ug/L
Fenuron	0.032	U	ug/L
Fluridone	8.0E-04	U	ug/L
Hydrocodone	0.0020	U	ug/L
Ibuprofen	0.020	U	ug/L
Imazapyr	0.0040	U	ug/L
Imidacloprid	0.0020	U	ug/L
Linuron	0.0040	U	ug/L
Mandestrobin	0.0020	U	ug/L
MCPP	0.0020	U	ug/L
Naproxen	0.0080	U	ug/L
Primidone	0.0040	U	ug/L
Pyraclostrobin	0.0020	U	ug/L
Silvex	0.0040	U	ug/L
Thiamethoxam	0.0020	U	ug/L
Tolfenpyrad	0.0020	U	ug/L
Triclopyr	0.0040	U	ug/L

Reference Method: EPA 8321B

Batch ID: P412593

Component	Result	Code	Units
Acesulfame K	0.10	U	ug/L
AMPA	0.10	U	ug/L
Endothall	0.25	U	ug/L
Glufosinate	0.10	U	ug/L
Glyphosate	0.10	U	ug/L
Sucralose	0.010	U	ug/L

Reference Method: SM 2320 B-2011

Batch ID: P413141

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO3/L

Reference Method: SM 2540 C-2011

Batch ID: P413063

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-2011

Batch ID: P412990

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 2540 D-2011

Batch ID: P412991

Component	Result	Code	Units
TSS	2	U	mg/L

Quality Assurance Report

Method Blank Results

Reference Method: SM 4500-F- C-2011
Batch ID: P413148

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-2011
Batch ID: P412742

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: DEP SOP: NU-094-1
Batch ID: P412564

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Color (true)	102		P	85 - 115

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P412573

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Turbidity	95.7		P	90 - 110

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P412588

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Calcium	102		P	85 - 115
Iron	105		P	85 - 115
Magnesium	107		P	85 - 115
Potassium	101		P	85 - 115
Sodium	104		P	85 - 115
Strontium	105		P	85 - 115
Tin	98.0		P	85 - 115
Titanium	97.2		P	85 - 115
Vanadium	93.3		P	85 - 115
Zinc	100		P	85 - 115

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P412588

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Aluminum	102		P	85 - 115
Antimony	102		P	85 - 115
Arsenic	102		P	85 - 115
Barium	105		P	85 - 115
Beryllium	100		P	85 - 115
Boron	99.8		P	85 - 115
Cadmium	99.2		P	85 - 115
Chromium	99.9		P	85 - 115
Cobalt	103		P	85 - 115
Copper	99.9		P	85 - 115
Lead	99.6		P	85 - 115
Manganese	101		P	85 - 115
Molybdenum	100		P	85 - 115
Nickel	101		P	85 - 115
Selenium	99.9		P	85 - 115
Silver	104		P	85 - 115
Thallium	103		P	85 - 115

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P412954

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	101		P	90 - 110

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P412957

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	100		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P412909

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	103		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P413093

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	97.2		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P412789

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	101		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P412790

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	100		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P412799

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	103		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P412555

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	98.2		P	90 - 110

Reference Method: EPA 8321B
Batch ID: P412554

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2,4-D	90.5		P	30 - 160
2,4,5-T	105		P	30 - 160
Acetaminophen	88.1		P	30 - 150
Acetamiprid	104		P	30 - 160
Afidopyropen	72.8		P	30 - 160
Bentazon	59.1		P	30 - 150
Benzovindiflupyr	96.1		P	30 - 160
Carbamazepine	110		P	30 - 160
Clothianidin	108		P	30 - 160
Dinotefuran	104		P	30 - 160

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 8321B

Batch ID: P412554

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Diuron	98.4		P	30 - 160
Fenuron	103		P	30 - 160
Fluridone	95.6		P	30 - 160
Hydrocodone	101		P	30 - 160
Ibuprofen	80.8		P	30 - 160
Imazapyr	98.9		P	30 - 160
Imidacloprid	104		P	30 - 160
Linuron	82.4		P	30 - 160
Mandestrobin	104		P	30 - 160
MCPP	116		P	30 - 160
Naproxen	112		P	30 - 160
Primidone	108		P	30 - 160
Pyraclostrobin	101		P	30 - 160
Silvex	106		P	30 - 160
Thiamethoxam	104		P	30 - 160
Tolfenpyrad	63.4		P	30 - 160
Triclopyr	112		P	30 - 160

Reference Method: EPA 8321B

Batch ID: P412593

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Acesulfame K	103		P	40 - 150
AMPA	99.6		P	40 - 150
Endothall	94.6		P	40 - 150
Glufosinate	96.6		P	40 - 150
Glyphosate	92.9		P	40 - 150
Sucralose	92.3		P	40 - 150

Reference Method: SM 2320 B-2011

Batch ID: P413141

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	96.7		P	94 - 106

Reference Method: SM 4500-F- C-2011

Batch ID: P413148

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	99.1		P	93 - 104

Reference Method: SM 5310 B-2011

Batch ID: P412742

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	95.0		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 200.7 Rev. 4.4

Batch ID: P412588

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2319897	Calcium	97.8		P	80 - 120
2319897	Iron	105	106	P/P	80 - 120
2319897	Magnesium	104	104	P/P	80 - 120
2319897	Potassium	101	102	P/P	80 - 120
2319897	Sodium	103	102	P/P	80 - 120
2319897	Strontium	106	105	P/P	80 - 120
2319897	Tin	97.1	97.9	P/P	80 - 120
2319897	Titanium	97.7	99.0	P/P	80 - 120
2319897	Vanadium	94.1	95.1	P/P	80 - 120
2319897	Zinc	98.9	97.5	P/P	80 - 120
2320056	Calcium	99.6		P	80 - 120
2320056	Iron	105		P	80 - 120
2320056	Magnesium	107		P	80 - 120
2320056	Potassium	99.1		P	80 - 120
2320056	Sodium	101		P	80 - 120
2320056	Strontium	105		P	80 - 120
2320056	Tin	98.9		P	80 - 120
2320056	Titanium	99.9		P	80 - 120
2320056	Vanadium	96.4		P	80 - 120
2320056	Zinc	101		P	80 - 120

Reference Method: EPA 200.8 Rev. 5.4

Batch ID: P412588

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2319897	Aluminum	99.1	101	P/P	80 - 120
2319897	Antimony	102	104	P/P	80 - 120
2319897	Arsenic	101	102	P/P	80 - 120
2319897	Barium	104	105	P/P	80 - 120
2319897	Beryllium	99.3	102	P/P	80 - 120
2319897	Boron	98.7	101	P/P	80 - 120
2319897	Cadmium	100	104	P/P	80 - 120
2319897	Chromium	100	102	P/P	80 - 120
2319897	Cobalt	102	104	P/P	80 - 120
2319897	Copper	97.7	99.1	P/P	80 - 120
2319897	Lead	98.7	101	P/P	80 - 120
2319897	Manganese	100	103	P/P	80 - 120
2319897	Molybdenum	100	102	P/P	80 - 120
2319897	Nickel	98.5	100	P/P	80 - 120
2319897	Selenium	98.9	100	P/P	80 - 120
2319897	Silver	103	105	P/P	80 - 120
2319897	Thallium	103	106	P/P	80 - 120
2320056	Aluminum	102		P	80 - 120
2320056	Antimony	103		P	80 - 120
2320056	Arsenic	103		P	80 - 120
2320056	Barium	105		P	80 - 120
2320056	Beryllium	102		P	80 - 120
2320056	Boron	105		P	80 - 120
2320056	Cadmium	103		P	80 - 120
2320056	Chromium	102		P	80 - 120
2320056	Cobalt	104		P	80 - 120
2320056	Copper	103		P	80 - 120

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P412588

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2320056	Lead	102		P	80 - 120
2320056	Manganese	104		P	80 - 120
2320056	Molybdenum	102		P	80 - 120
2320056	Nickel	102		P	80 - 120
2320056	Selenium	99.9		P	80 - 120
2320056	Silver	103		P	80 - 120
2320056	Thallium	105		P	80 - 120

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P412954

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2320050	Chloride	101		P	90 - 110
2320486	Chloride	99.6		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P412957

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2320050	Sulfate	96.7		P	90 - 110
2320486	Sulfate	99.5		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P412909

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2320187	Ammonia-N	101	101	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P413093

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2320049	Kjeldahl Nitrogen	109	102	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P412789

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2319680	NO2NO3-N	99.8	99.8	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P412790

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2320364	NO2NO3-N	93.2		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P412799

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2320250	Total-P	102	103	P/P	90 - 110

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P412555

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2319893	O-Phosphate-P	97.9	99.9	P/P	90 - 110

Reference Method: EPA 8321B
Batch ID: P412554

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2319995	2,4-D	108	84.5	P/P	30 - 160
2319995	2,4,5-T	109	82.1	P/P	30 - 160
2319995	Acetaminophen	97.0	80.2	P/P	30 - 150
2319995	Acetamiprid	115	102	P/P	30 - 160
2319995	Afidopyropen	80.9	76.6	P/P	30 - 160
2319995	Bentazon	39.7	31.9	P/P	30 - 150
2319995	Benzovindiflupyr	88.7	76.5	P/P	30 - 160
2319995	Carbamazepine	102	88.3	P/P	30 - 160
2319995	Clothianidin	130	110	P/P	30 - 160
2319995	Dinotefuran	128	115	P/P	30 - 160
2319995	Diuron	84.0	73.4	P/P	30 - 160
2319995	Fenuron	87.8	77.7	P/P	30 - 160
2319995	Fluridone	90.0	75.0	P/P	30 - 160
2319995	Hydrocodone	110	89.0	P/P	30 - 160
2319995	Ibuprofen	76.8	68.1	P/P	30 - 160
2319995	Imazapyr	112	90.0	P/P	30 - 160
2319995	Imidacloprid	184	159	F/P	30 - 160
2319995	Linuron	70.8	64.4	P/P	30 - 160
2319995	Mandestrobin	95.0	81.5	P/P	30 - 160
2319995	MCPP	129	98.5	P/P	30 - 160
2319995	Naproxen	92.3	86.1	P/P	30 - 160
2319995	Primidone	106	105	P/P	30 - 160
2319995	Pyraclostrobin	97.7	80.9	P/P	30 - 160
2319995	Silvex	113	80.5	P/P	30 - 160
2319995	Thiamethoxam	170	148	F/P	30 - 160
2319995	Tolfenpyrad	61.0	48.9	P/P	30 - 160
2319995	Triclopyr	122	96.2	P/P	30 - 160

Reference Method: EPA 8321B
Batch ID: P412593

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2319878	Acesulfame K	106	110	P/P	40 - 150
2319878	AMPA	106	108	P/P	40 - 150
2319878	Endothall	100	104	P/P	40 - 150
2319878	Glufosinate	98.5	106	P/P	40 - 150
2319878	Glyphosate	89.8	90.1	P/P	40 - 150
2319878	Sucralose	99.0	102	P/P	40 - 150

Reference Method: SM 4500-F- C-2011
Batch ID: P413148

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2320031	Fluoride	98.6	98.6	P/P	92 - 107

Quality Assurance Report

Matrix Spike Accuracy

Reference Method: SM 5310 B-2011
Batch ID: P412742

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2320074	Organic Carbon	99.7	98.4	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: DEP SOP: NU-094-1
Batch ID: P412564

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2320073	Color (true)	0.712	Sample	P	0 - 20

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P412573

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2320050	Turbidity	11.6	Sample	P	0 - 20

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P412588

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2319897	Calcium	1.33	Spike	P	0 - 20
2319897	Iron	0.711	Spike	P	0 - 20
2319897	Magnesium	0.124	Spike	P	0 - 20
2319897	Potassium	0.912	Spike	P	0 - 20
2319897	Sodium	0.696	Spike	P	0 - 20
2319897	Strontium	1.15	Spike	P	0 - 20
2319897	Tin	0.828	Spike	P	0 - 20
2319897	Titanium	1.31	Spike	P	0 - 20
2319897	Vanadium	1.05	Spike	P	0 - 20
2319897	Zinc	1.42	Spike	P	0 - 20

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P412588

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2319897	Aluminum	1.64	Spike	P	0 - 20
2319897	Antimony	1.83	Spike	P	0 - 20
2319897	Arsenic	1.67	Spike	P	0 - 20
2319897	Barium	1.06	Spike	P	0 - 20
2319897	Beryllium	2.70	Spike	P	0 - 20
2319897	Boron	1.97	Spike	P	0 - 20
2319897	Cadmium	3.23	Spike	P	0 - 20
2319897	Chromium	1.30	Spike	P	0 - 20
2319897	Cobalt	1.61	Spike	P	0 - 20
2319897	Copper	1.46	Spike	P	0 - 20
2319897	Lead	2.21	Spike	P	0 - 20
2319897	Manganese	2.23	Spike	P	0 - 20
2319897	Molybdenum	2.14	Spike	P	0 - 20
2319897	Nickel	2.00	Spike	P	0 - 20
2319897	Selenium	1.52	Spike	P	0 - 20
2319897	Silver	2.52	Spike	P	0 - 20
2319897	Thallium	2.34	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P412954

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2320050	Chloride	0.403	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P412957

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2320050	Sulfate	0.497	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P412909

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2320187	Ammonia-N	0.151	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P413093

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2320049	Kjeldahl Nitrogen	4.42	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P412789

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2319680	NO2NO3-N	0.0	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P412790

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2320364	NO2NO3-N	0.177	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P412799

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2320250	Total-P	1.06	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P412555

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2319893	O-Phosphate-P	1.65	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: EPA 8321B
Batch ID: P412554

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2319995	2,4-D	24.1	Spike	P	0 - 30
2319995	2,4,5-T	28.6	Spike	P	0 - 30
2319995	Acetaminophen	18.9	Spike	P	0 - 30
2319995	Acetamiprid	11.4	Spike	P	0 - 30
2319995	Afidopyropen	5.51	Spike	P	0 - 30
2319995	Bentazon	21.7	Spike	P	0 - 30
2319995	Benzovindiflupyr	14.8	Spike	P	0 - 30
2319995	Carbamazepine	14.6	Spike	P	0 - 30
2319995	Clothianidin	17.0	Spike	P	0 - 30
2319995	Dinotefuran	10.6	Spike	P	0 - 30
2319995	Diuron	13.4	Spike	P	0 - 30
2319995	Fenuron	12.1	Spike	P	0 - 30
2319995	Fluridone	18.2	Spike	P	0 - 30
2319995	Hydrocodone	21.0	Spike	P	0 - 30
2319995	Ibuprofen	12.0	Spike	P	0 - 30
2319995	Imazapyr	21.5	Spike	P	0 - 30
2319995	Imidacloprid	14.8	Spike	P	0 - 30
2319995	Linuron	9.42	Spike	P	0 - 30
2319995	Mandestrobin	15.3	Spike	P	0 - 30
2319995	MCP	26.9	Spike	P	0 - 30
2319995	Naproxen	6.90	Spike	P	0 - 30
2319995	Primidone	0.363	Spike	P	0 - 30
2319995	Pyraclostrobin	18.9	Spike	P	0 - 30
2319995	Silvex	33.3	Spike	F	0 - 30
2319995	Thiamethoxam	13.5	Spike	P	0 - 30
2319995	Tolfenpyrad	22.1	Spike	P	0 - 30
2319995	Triclopyr	23.3	Spike	P	0 - 30

Reference Method: EPA 8321B
Batch ID: P412593

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2319878	Acesulfame K	3.55	Spike	P	0 - 30
2319878	AMPA	2.21	Spike	P	0 - 30
2319878	Endothall	3.93	Spike	P	0 - 30
2319878	Glufosinate	7.61	Spike	P	0 - 30
2319878	Glyphosate	0.343	Spike	P	0 - 30
2319878	Sucralose	3.20	Spike	P	0 - 30

Reference Method: SM 2320 B-2011
Batch ID: P413141

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2320031	Total Alkalinity	0.296	Sample	P	0 - 5.2

Quality Assurance Report Precision

Reference Method: SM 2540 C-2011
Batch ID: P413063

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2320032	TDS	2.11	Sample	P	0 - 10

Reference Method: SM 4500-F- C-2011
Batch ID: P413148

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2320031	Fluoride	0.0	Spike	P	0 - 2.1

Reference Method: SM 5310 B-2011
Batch ID: P412742

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2320074	Organic Carbon	0.755	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.

Quality Assurance Report Surrogates

Lab Sample ID: 2320197

Field Sample ID: G4NW0080

Reference Method	Surrogate	% Rec.	Pass/Fail	Control Limits
EPA 8321B	2,4-D-d3	81.1	P	30 - 160
EPA 8321B	Acetaminophen-d4	85.4	P	30 - 160
EPA 8321B	Carbamazepine-d10	83.0	P	30 - 160
EPA 8321B	Diuron-d6	63.8	P	30 - 160
EPA 8321B	Endothall-d6	99.7	P	30 - 160
EPA 8321B	Glyphosate 13C 15N	76.7	P	30 - 160
EPA 8321B	Ibuprofen-d3	56.7	P	30 - 160
EPA 8321B	Primidone-d5	76.5	P	30 - 160
EPA 8321B	Sucralose-d6	101	P	30 - 160

Quality Assurance Report Calibration Verification

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A111554

Included Lab Sample IDs: 2320188, 2320191

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	98.8	99.5	P/P	90 - 110

Reference Method: DEP SOP: NU-094-1

Run ID: A111556

Included Lab Sample IDs: 2320189

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	103	104	P/P	85 - 115

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A111559

Included Lab Sample IDs: 2320186, 2320189

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	99.5	99.0	P/P	95 - 105

Reference Method: EPA 200.8 Rev. 5.4

Run ID: A111617

Included Lab Sample IDs: 2320198

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Aluminum	102	101	P/P	90 - 110
Antimony	103	103	P/P	90 - 110
Arsenic	102	101	P/P	90 - 110
Barium	100	101	P/P	90 - 110
Beryllium	98.2	98.6	P/P	90 - 110
Boron	98.5	97.8	P/P	90 - 110
Cadmium	102	101	P/P	90 - 110
Chromium	99.9	99.7	P/P	90 - 110
Cobalt	103	103	P/P	90 - 110
Copper	102	101	P/P	90 - 110
Lead	98.5	98.4	P/P	90 - 110
Manganese	102	102	P/P	90 - 110
Molybdenum	99.6	99.6	P/P	90 - 110
Nickel	102	102	P/P	90 - 110
Selenium	100	101	P/P	90 - 110
Silver	105	105	P/P	90 - 110
Thallium	98.8	98.1	P/P	90 - 110

Reference Method: SM 5310 B-2011

Run ID: A111622

Included Lab Sample IDs: 2320187, 2320190

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	97.2	97.8	P/P	90 - 110
Organic Carbon	97.8	98.0	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 8321B
Run ID: A111640
Included Lab Sample IDs: 2320197

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Acesulfame K	95.4	92.3	P/P	60 - 160
Sucralose	99.3	99.5	P/P	60 - 160

Reference Method: EPA 353.2 Rev. 2.0
Run ID: A111649
Included Lab Sample IDs: 2320187, 2320190

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	100	100	P/P	90 - 110
NO2NO3-N	102	100	P/P	90 - 110

Reference Method: EPA 8321B
Run ID: A111655
Included Lab Sample IDs: 2320197

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
AMPA	106	110	P/P	60 - 160
Endothall	100	102	P/P	60 - 160
Glufosinate	106	106	P/P	60 - 160
Glyphosate	104	103	P/P	60 - 160

Reference Method: EPA 200.7 Rev. 4.4
Run ID: A111687
Included Lab Sample IDs: 2320198

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Calcium	98.2	99.5	P/P	90 - 110
Iron	99.3	101	P/P	90 - 110
Magnesium	99.6	101	P/P	90 - 110
Potassium	100	102	P/P	90 - 110
Sodium	99.3	100	P/P	90 - 110
Strontium	108	107	P/P	90 - 110
Tin	99.3	99.6	P/P	90 - 110
Titanium	98.6	99.6	P/P	90 - 110
Vanadium	99.7	101	P/P	90 - 110
Zinc	101	102	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Run ID: A111696
Included Lab Sample IDs: 2320187, 2320190

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	100	98.5	P/P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Run ID: A111715
Included Lab Sample IDs: 2320189

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	101	101	P/P	90 - 110
Sulfate	101	101	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 8321B
Run ID: A111767
Included Lab Sample IDs: 2320197

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
2,4-D	95.7	84.7	P/P	50 - 150
2,4,5-T	105	89.4	P/P	50 - 150
Acetaminophen	99.4	86.3	P/P	60 - 160
Acetamiprid	105	95.6	P/P	50 - 150
Afidopyropen	95.7	83.3	P/P	50 - 150
Bentazon	106	80.5	P/P	50 - 160
Benzovindiflupyr	113	92.6	P/P	50 - 150
Carbamazepine	105	93.3	P/P	60 - 150
Clothianidin	105	90.7	P/P	60 - 150
Dinotefuran	103	94.1	P/P	50 - 150
Diuron	105	97.2	P/P	60 - 160
Fenuron	101	97.6	P/P	60 - 150
Fluridone	107	107	P/P	60 - 160
Hydrocodone	104	87.6	P/P	60 - 150
Ibuprofen	107	110	P/P	50 - 160
Imazapyr	87.7	85.1	P/P	50 - 150
Imidacloprid	101	89.6	P/P	60 - 150
Linuron	98.3	95.8	P/P	60 - 150
Mandestrobin	112	103	P/P	50 - 150
MCPP	84.6	73.7	P/P	50 - 150
Naproxen	104	72.6	P/P	50 - 160
Primidone	91.2	91.9	P/P	60 - 150
Pyraclostrobin	120	108	P/P	60 - 150
Silvex	88.4	78.4	P/P	50 - 150
Thiamethoxam	102	91.9	P/P	50 - 150
Tolfenpyrad	108	99.4	P/P	60 - 150
Triclopyr	110	85.1	P/P	50 - 150

Reference Method: EPA 365.1 Rev. 2.0
Run ID: A111774
Included Lab Sample IDs: 2320187, 2320190

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	103	104	P/P	90 - 110
Total-P	104	103	P/P	90 - 110

Reference Method: SM 2320 B-2011
Run ID: A111807
Included Lab Sample IDs: 2320186, 2320189

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	96.3	94.4	P/P	90 - 110

Reference Method: SM 4500-F- C-2011
Run ID: A111807
Included Lab Sample IDs: 2320186, 2320189

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	99.0	99.5	P/P	90 - 110

Quality Assurance Report

Calibration Verification

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A111862

Included Lab Sample IDs: 2320187, 2320190

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	102	92.6	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.

Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery			LCS	Precision SMP	MS
DEP SOP: NU-094-1	Color (true)	102					0.712	
EPA 180.1 Rev. 2.0	Turbidity	95.7					11.6	
EPA 200.7 Rev. 4.4	Calcium	102	97.8	99.6				1.33
	Iron	105	105	106	105			0.711
	Magnesium	107	104	104	107			0.124
	Potassium	101	101	102	99.1			0.912
	Sodium	104	103	102	101			0.696
	Strontium	105	106	105	105			1.15
	Tin	98.0	97.1	97.9	98.9			0.828
	Titanium	97.2	97.7	99.0	99.9			1.31
	Vanadium	93.3	94.1	95.1	96.4			1.05
	Zinc	100	98.9	97.5	101			1.42
EPA 200.8 Rev. 5.4	Aluminum	102	99.1	101	102			1.64
	Antimony	102	102	104	103			1.83
	Arsenic	102	101	102	103			1.67
	Barium	105	104	105	105			1.06
	Beryllium	100	99.3	102	102			2.70
	Boron	99.8	98.7	101	105			1.97
	Cadmium	99.2	100	104	103			3.23
	Chromium	99.9	100	102	102			1.30
	Cobalt	103	102	104	104			1.61
	Copper	99.9	97.7	99.1	103			1.46
	Lead	99.6	98.7	101	102			2.21
	Manganese	101	100	103	104			2.23
	Molybdenum	100	100	102	102			2.14
	Nickel	101	98.5	100	102			2.00
	Selenium	99.9	98.9	100	99.9			1.52
	Silver	104	103	105	103			2.52
	Thallium	103	103	106	105			2.34
EPA 300.0 Rev. 2.1	Chloride	101	101	99.6			0.403	
	Sulfate	100	96.7	99.5			0.497	
EPA 350.1 Rev. 2.0	Ammonia-N	103	101	101				0.151
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	97.2	109	102				4.42
EPA 353.2 Rev. 2.0	NO2NO3-N	101	99.8	99.8				0.0
	NO2NO3-N	100	93.2					0.177
EPA 365.1 Rev. 2.0	Total-P	103	102	103				1.06
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	98.2	97.9	99.9				1.65
EPA 8321B	2,4-D	90.5	108	84.5				24.1
	2,4,5-T	105	109	82.1				28.6
	Acesulfame K	103	106	110				3.55
	Acetaminophen	88.1	97.0	80.2				18.9
	Acetamiprid	104	115	102				11.4
	Afidopyropen	72.8	80.9	76.6				5.51
	AMPA	99.6	106	108				2.21
	Bentazon	59.1	39.7	31.9				21.7
	Benzovindiflupyr	96.1	88.7	76.5				14.8
	Carbamazepine	110	102	88.3				14.6
	Clothianidin	108	130	110				17.0
	Dinotefuran	104	128	115				10.6
	Diuron	98.4	84.0	73.4				13.4
	Endothall	94.6	100	104				3.93
	Fenuron	103	87.8	77.7				12.1
	Fluridone	95.6	90.0	75.0				18.2
	Glufosinate	96.6	98.5	106				7.61

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		Precision LCS	SMP	MS
EPA 8321B	Glyphosate	92.9	89.8	90.1			0.343
	Hydrocodone	101	110	89.0			21.0
	Ibuprofen	80.8	76.8	68.1			12.0
	Imazapyr	98.9	112	90.0			21.5
	Imidacloprid	104	184	159			14.8
	Linuron	82.4	70.8	64.4			9.42
	Mandestrobin	104	95.0	81.5			15.3
	MCPP	116	129	98.5			26.9
	Naproxen	112	92.3	86.1			6.90
	Primidone	108	106	105			0.363
	Pyraclostrobin	101	97.7	80.9			18.9
	Silvex	106	113	80.5			33.3
	Sucralose	92.3	99.0	102			3.20
	Thiamethoxam	104	170	148			13.5
	Tolfenpyrad	63.4	61.0	48.9			22.1
	Triclopyr	112	122	96.2			23.3
SM 2320 B-2011	Total Alkalinity	96.7				0.296	
SM 2540 C-2011	TDS					2.11	
SM 4500-F- C-2011	Fluoride	99.1	98.6	98.6			0.0
SM 5310 B-2011	Organic Carbon	95.0	99.7	98.4			0.755

Reference Method Descriptions

Method	Description	Associated Samples
DEP SOP: NU-094-1	True Color measured at 450 nm (not valid for NPDES monitoring)	2320189
EPA 180.1 Rev. 2.0	Turbidity in aqueous matrices	2320186, 2320189
EPA 200.7 Rev. 4.4	Total Recoverable Metals analysis using ICP emission spectroscopy for aqueous samples supporting Clean Water Act Projects	2320198
EPA 200.8 Rev. 5.4	Total Recoverable Metals analysis using ICP-MS for aqueous samples supporting Clean Water Act Projects	2320198
EPA 300.0 Rev. 2.1	Chloride in aqueous matrices	2320189
EPA 300.0 Rev. 2.1	Sulfate in aqueous matrices	2320189
EPA 350.1 Rev. 2.0	Ammonia in aqueous matrices as mg N/L	2320187, 2320190
EPA 351.2 Rev. 2.0	Total Kjeldahl Nitrogen in aqueous matrices	2320187, 2320190
EPA 353.2 Rev. 2.0	Nitrite/Nitrate in aqueous matrices as mg N/L	2320187, 2320190
EPA 365.1 Rev. 2.0	Total Phosphorus in aqueous matrices as mg P/L	2320187, 2320190
EPA 365.1 Rev. 2.0 dissolved	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	2320188, 2320191
EPA 8321B	Pesticides, herbicides, and other chemicals in water matrices by HPLC/MS/MS	2320197
SM 2320 B-2011	Alkalinity in aqueous matrices as mg CaCO3/L	2320186, 2320189
SM 2540 C-2011	Total Dissolved Solids in aqueous matrices	2320189
SM 2540 D-2011	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	2320186, 2320189
SM 4500-F- C-2011	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	2320186, 2320189
SM 5310 B-2011	Nonpurgeable Organic Carbon in aqueous matrices	2320187, 2320190

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
DEP SOP: NU-094-1	04/19/2022			04/19/2022 15:51	Anna M. Plaviak	2320189
EPA 180.1 Rev. 2.0	04/19/2022			04/19/2022 14:46	Sarah A Nixon	2320189
	04/19/2022			04/19/2022 14:52	Sarah A Nixon	2320186
EPA 200.7 Rev. 4.4	04/19/2022	04/20/2022 10:20	Megan Whitefoot	04/25/2022 17:44	McKinley C Carter	2320198
EPA 200.8 Rev. 5.4	04/19/2022	04/20/2022 10:20	Megan Whitefoot	04/21/2022 23:46	Alexander Thompson	2320198
EPA 300.0 Rev. 2.1	04/19/2022			04/27/2022 02:03	Anna M. Plaviak	2320189
EPA 350.1 Rev. 2.0	04/19/2022			04/26/2022 14:54	Judith K. Clark	2320187
	04/19/2022			04/26/2022 15:08	Judith K. Clark	2320190
EPA 351.2 Rev. 2.0	04/19/2022	04/29/2022 09:35	Samantha L Bates	05/02/2022 15:38	Alexandra J Mattheus	2320187
	04/19/2022	04/29/2022 09:35	Samantha L Bates	05/02/2022 15:40	Alexandra J Mattheus	2320190
EPA 353.2 Rev. 2.0	04/19/2022			04/22/2022 15:15	Nathaniel J Jones	2320187
	04/19/2022			04/22/2022 15:35	Nathaniel J Jones	2320190
EPA 365.1 Rev. 2.0	04/19/2022	04/26/2022 14:30	Simonne.V. Calhoun	04/28/2022 14:27	James L Waggeberby	2320187
	04/19/2022	04/26/2022 14:30	Simonne.V. Calhoun	04/28/2022 14:35	James L Waggeberby	2320190
EPA 365.1 Rev. 2.0 dissolved	04/19/2022			04/19/2022 14:27	James L Waggeberby	2320188
	04/19/2022			04/19/2022 14:28	James L Waggeberby	2320191
EPA 8321B	04/19/2022	04/20/2022 09:00	Hoor Shaik	04/21/2022 11:14	Juyoung Kim	2320197
	04/19/2022	04/20/2022 10:50	Manjita Shrestha	04/20/2022 21:27	Manjita Shrestha	2320197
	04/19/2022	04/20/2022 10:50	Manjita Shrestha	04/22/2022 16:09	Manjita Shrestha	2320197
SM 2320 B-2011	04/19/2022			04/28/2022 21:31	Dale L. Simmons	2320186
	04/19/2022			04/28/2022 21:41	Dale L. Simmons	2320189
SM 2540 C-2011	04/19/2022			04/20/2022 15:36	Yijie Li	2320189
SM 2540 D-2011	04/19/2022			04/20/2022 15:36	Yijie Li	2320186, 2320189
SM 4500-F- C-2011	04/19/2022			04/28/2022 21:31	Dale L. Simmons	2320186
	04/19/2022			04/28/2022 21:41	Dale L. Simmons	2320189
SM 5310 B-2011	04/19/2022			04/21/2022 17:29	Khloe J Berg	2320187
	04/19/2022			04/21/2022 21:59	Khloe J Berg	2320190