


BENCHMARK

EnviroAnalytical Inc.



NELAC Certification #E84167

ANALYTICAL TEST REPORT

THESE RESULTS MEET NELAC STANDARDS

Submission Number : 22040080

F D E P/S W D
13051 N Telecom Pkwy.
Tampa, FL 33637

Project Name : RQ-2022-04-04-16
Date Received : 04/04/2022
Time Received : 1400

Erica Peck

Submission Number: 22040080 Sample Date: 04/04/2022
Sample Number: 001 Sample Time: 1040
Sample Description: G1SW0139 Sample Method: Grab

Parameter	Result	Units	MDL	PQL	Procedure	Preparation Date/Time	Analysis Date/Time	Analyst
ENTEROCOCCI	158	#/100 ML	10	10	ENTEROLERT	04/04/2022 16:20	04/05/2022 16:34	LG

Submission Number: 22040080 Sample Date: 04/04/2022
Sample Number: 002 Sample Time: 1045
Sample Description: G1SW0102 Sample Method: Grab

Parameter	Result	Units	MDL	PQL	Procedure	Preparation Date/Time	Analysis Date/Time	Analyst
ENTEROCOCCI	749	#/100 ML	10	10	ENTEROLERT	04/04/2022 16:20	04/05/2022 16:34	LG

Submission Number: 22040080 Sample Date: 04/04/2022
Sample Number: 003 Sample Time: 1130
Sample Description: G1SW0112 Sample Method: Grab


Parameter	Result	Units	MDL	PQL	Procedure	Preparation Date/Time	Analysis Date/Time	Analyst
ENTEROCOCCI	560	#/100 ML	10	10	ENTEROLERT	04/04/2022 16:20	04/05/2022 16:34	LG

Submission Number: 22040080 Sample Date: 04/04/2022
Sample Number: 004 Sample Time: 1215
Sample Description: G1SW0122 Sample Method: Grab

Parameter	Result	Units	MDL	PQL	Procedure	Preparation Date/Time	Analysis Date/Time	Analyst
E- COLI BY IDEXX QUANTITRAY	214	#/100 ML	10	10	SM9223B	04/04/2022 16:35	04/05/2022 16:45	LG

BENCHMARK

EnviroAnalytical Inc.



NELAC Certification #E84167

Submission Number: 22040080
Sample Number: 005
Sample Description: G1SW0121


Sample Date: 04/04/2022
Sample Time: 1240
Sample Method: Grab

Parameter	Result	Units	MDL	PQL	Procedure	Preparation Date/Time	Analysis Date/Time	Analyst
ENTEROCOCCI	1850	#/100 ML	10	10	ENTEROLERT	04/04/2022 16:20	04/05/2022 16:34	LG
FECAL COLIFORM	270	#/100 ML	10	10	SM9222D	04/04/2022 16:50	04/05/2022 15:10	LG
E- COLI BY IDEXX QUANTITRAY	233	#/100 ML	10	10	SM9223B	04/04/2022 16:35	04/05/2022 16:45	LG

Submission Number: 22040080
Sample Number: 006
Sample Description: G1SW0120

Sample Date: 04/04/2022
Sample Time: 1310
Sample Method: Grab

Parameter	Result	Units	MDL	PQL	Procedure	Preparation Date/Time	Analysis Date/Time	Analyst
ENTEROCOCCI	20	#/100 ML	10	10	ENTEROLERT	04/04/2022 16:20	04/05/2022 16:34	LG
FECAL COLIFORM	20	#/100 ML	1	1	SM9222D	04/04/2022 16:50	04/05/2022 15:10	LG
E- COLI BY IDEXX QUANTITRAY	345	#/100 ML	10	10	SM9223B	04/04/2022 16:35	04/05/2022 16:45	LG


Dale D. Dixon / Laboratory Director

04/07/2022

Date

Tülay Tanrısever - Technical Director/QC Officer

Kara Peterson - QC Officer

DATA QUALIFIERS THAT MAY APPLY:

A = Value reported is an average of two or more determinations.
B = Results based upon colony counts outside the ideal range.
H = Value based on field kit determination. Results may not be accurate.
I = Reported value is between the laboratory MDL and the PQL.
J1 = Estimated value. Surrogate recovery limits exceeded.
J2 = Estimated value. No quality control criteria exists for component.
J3 = Estimated value. Quality control criteria for precision or accuracy not met.
J4 = Estimated value. Sample matrix interference suspected.
J5 = Estimated value. Data questionable due to improper lab or field protocols.
K = Off-scale low. Value is known to be < the value reported.
L = Off-scale high. Value is known to be > the value reported.
N = Presumptive evidence of presence of material.
O = Sampled, but analysis lost or not performed.

Q = Sample held beyond accepted hold time.
T = Value reported is < MDL. Reported for informational purposes only and shall not be used in statistical analysis.
U = Analyte analyzed but not detected at the value indicated.
V = Analyte detected in sample and method blank. Results for this analyte in associated samples may be biased high. Standard, Duplicate and Spike values are within control limits. Reported data are usable.
Y = Analysis performed on an improperly preserved sample. Data may be inaccurate.
Z = Too many colonies were present (TNTC). The numeric value represents the filtration volume.
I = Data deviate from historically established concentration ranges.
? = Data rejected and should not be used. Some or all of QC data were outside criteria, and the presence or absence of the analyte cannot be determined from the data.
* = Not reported due to interference.
Oil & Grease - If client does not send sufficient sample quantity for spike evaluation surface water samples are supplied by the laboratory.

NOTES:

MBAS calculated as LAS; molecular weight = 340.
PQL = 4xMDL.
ND = Not detected at or above the adjusted reporting limit.
X = Value exceeds MCL.
G1 = Accuracy standard does not meet method control limits, but does meet lab control limits that are in agreement with USEPA generated data. USEPA letter available upon request.

COMMENTS:

1 (10 mL) dilution on each Enterococci and E. coli sample. 3 (1mL, 10 mL, 100 mL) dilutions on each Fecal Coliform sample.

For questions or comments regarding these results, please contact us at (941) 723-9986.

Results relate only to the samples.

Benchmark EnviroAnalytical, Inc
1711 12th Street East
Palmetto, FL 34221
941-723-9986
941-723-6061 Fax

Client Information: DEP: SW-DIST

Address: 13051 N. Telecom Parkway
Tampa FL 33604
Phone # 815-470-5878

PO# B77880

Report to: benjamin.paswater@dep.state.fl.us
DEP-Overflomicro@dep.state.fl.us

Invoice to: FDEP Micro Overflow

Project Name: RQ-2022-04-04-16 **Laboratory Submission # 2204080**

Sample Name	Sample Type ¹ / Sample Matrix ²	Collection		Container		Preservative ⁴	Parameters for Analysis	Laboratory Sample #
		Date	Time	Qty	Capacity			
G1SW0139	SW	4/4/22	1040	1	250mL	P	• Entero	1
G1SW0102	SW		1045	1	250mL	P	• Entero	2
G1SW0112	SW		1130	1	250mL	P	• Entero	3
G1SW0122	SW		1215	1	250mL	P	• E. Coli	4
G1SW0121	SW		1240	3	250mL	P	• Entero, E. Coli, Fecal	5
G1SW0120	SW		1310	3	250mL	P	• Entero, E. Coli, Fecal	6
	SW				250mL	P		

¹ "Sample Type" is used to indicate whether the sample was a grab (G) or whether it was a composite (C).
² "Sample Matrix" is used to indicate whether the sample is being discharged to drinking water (DW), groundwater (GW), surface water (SW), fresh surface water (FSW), saline surface water (SSW), soil, sediment (SDSNT), or sludge (SLDS).
³ "Sample must be refrigerated or stored in wet ice after collection. The temperature during storage should be less than or equal to 6°C (42.8°F).
⁴ Under "Preservative," list any preservatives that were added to the sample container.

Instructions:
1. The client has a label identifying sample ID, preservative, and parameters for analysis.
2. The following information should be added to each bottle label after collection with permanent black ink: date and time of collection, sample's name or initials, and any field number or ID.
3. All bottles not containing preservative may be rinsed with appropriate sample prior to collection.
4. The client is responsible for documentation of the sampling event. Please note special sampling events on the sample custody form.

Laboratory Sample Acceptability:
Temperature: 22°C

1	Collected / Relinquished By: <i>Benjamin Paswater</i>	Date: 4/4/22	Time: 1400	Received By: <i>Benjamin Paswater</i>	Date: 4/4/22	Time: 1400
		Date:	Time:		Date:	Time:
2	Relinquished By:	Date:	Time:	Received By:	Date:	Time: