

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there is qualified data on this page.

Meter ID: BIG 7 RQ: 2018-01-29-67 Project: CONT MONIT. SMP

IMPERIV6R MARINE

Notes: (1) Always wait for meter to stabilize before recording any readings.

(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).

(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Circle/Fill In: Lab Calibration or Calibration/Verification on Site

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 1/25/18

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	JOHN BARRINGTON	1/25/18	0737	25.0	8.3	8.3	100	N/A	1.1	(P) F	(L) F
ICV	JOHN BARRINGTON	1/25/18	0750	25.4	8.2	8.4	100	—	—	(P) F	(L) F
CCV	JOHN BARRINGTON	1/30/18	1545	25.5	8.2	8.5	106	—	—	(P) F	(L) F
CCV										P / F	L / F

DO Acceptance criteria from Table ± 0.3 mg/L.

Optical: DO gain range 0.85 to 1.15; DO charge N/A.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ hos/cm	Meter Reading μ hos/cm	Pass / Fail	Lab / Field
Calibr.	JOHN BARRINGTON	0755	1/25/18	171211C	1/18	1000	1000	(P) F	(L) F
ICV	JOHN BARRINGTON	0758	1/25/18	170421A	2/5/18	10,000	9990	(P) F	(L) F
CCV	Greg Smith	1/30/18	1320	2706591	05/19	50,000	49589	(P) F	(L) F
CCV								P / F	L / F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	JOHN BARRINGTON	1/25/18	0800	176157	8/19	7.0	7.0	-25.0	(P) F	(L) F
Calibr.	JOHN BARRINGTON	1/25/18	0803	176044	8/19	4.0	4.0	152.1	(P) F	(L) F
Calibr.	JOHN BARRINGTON	1/25/18	0806	172902	4/19	10.0	10.0	-200	(P) F	(L) F
ICV	JOHN BARRINGTON	1/25/18	0809	176044	8/19	4.0	3.9	158.3	(P) F	(L) F
CCV	Greg Smith	1/30/18	1330	172902	4/19	10.0	10.0	-201.9	(P) F	(L) F
CCV									P / F	L / F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 175, slope from 4 to 7 177.1 (both must be between 165 and 180 mV)

Is meter equipped with depth sensor? Yes / No If yes, list date of last quarterly depth verification

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value meters	Pass / Fail	Lab / Field
Pressure mode in air						P / F	L / F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there is qualified data on this page.

Meter ID: B: 17

RQ: 2018-01-29-67

Project: Continuous monitoring SAP

SILVER CREEK

Notes: (1) Always wait for meter to stabilize before recording any readings.

(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).

(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Circle/Fill In: Lab Calibration or Calibration/Verification on Site

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 1/25/18

DO FDEP SOP FT 1500	Name	Date	Time CT-ET	Temp	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	<u>Gregory Sel</u>	<u>1/25/18</u>	<u>900</u>	<u>25.9</u>	<u>8.13</u>	<u>8.1</u>	<u>100</u>	<u>NA</u>	<u>NA</u>	<u>(P) F</u>	<u>(L) F</u>
ICV	<u>Gregory Sel</u>	<u>1/25/18</u>	<u>905</u>	<u>26.1</u>	<u>8.1</u>	<u>8.1</u>	<u>99</u>	<u>NA</u>	<u>1.07</u>	<u>(P) F</u>	<u>(L) F</u>
CCV	<u>John Barrington</u>	<u>1/30/18</u>	<u>1300</u>	<u>25.0</u>	<u>8.3</u>	<u>8.2</u>	<u>98.4</u>	<u>NA</u>	<u>1.07</u>	<u>(P) F</u>	<u>(L) F</u>
CCV										<u>P / F</u>	<u>L / F</u>

DO Acceptance criteria from Table ± 0.3 mg/L.

Optical: DO gain range 0.85 to 1.15; DO charge N/A.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ hos/cm	Meter Reading μ hos/cm	Pass / Fail	Lab / Field
Calibr.	<u>Gregory Sel</u>	<u>1/25/18</u>	<u>908</u>	<u>171211C</u>	<u>12/18</u>	<u>1000</u>	<u>1000</u>	<u>(P) F</u>	<u>(L) F</u>
ICV	<u>"</u>	<u>"</u>	<u>912</u>	<u>170728C</u>	<u>8/18</u>	<u>10,000</u>	<u>9590</u>	<u>(P) F</u>	<u>(L) F</u>
CCV	<u>John Barrington</u>	<u>1/30/18</u>	<u>1304</u>	<u>2706591</u>	<u>5/19</u>	<u>50,000</u>	<u>47484</u>	<u>(P) F</u>	<u>(L) F</u>
CCV	<u>John Barrington</u>	<u>1/30/18</u>	<u>1306</u>	<u>171211C</u>	<u>12/18</u>	<u>1000</u>	<u>1006</u>	<u>(P) F</u>	<u>(L) F</u>

Conductivity Acceptance criteria $\pm 5\%$

pH FDEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	<u>Gregory Sel</u>	<u>1/25/18</u>	<u>915</u>	<u>176157</u>	<u>8/19</u>	<u>7.0</u>	<u>7.0</u>	<u>-32.0</u>	<u>(P) F</u>	<u>(L) F</u>
Calibr.	<u>"</u>	<u>1/25/18</u>	<u>920</u>	<u>176044</u>	<u>8/19</u>	<u>4.0</u>	<u>4.0</u>	<u>131.0</u>	<u>(P) F</u>	<u>(L) F</u>
Calibr.	<u>"</u>	<u>1/25/18</u>	<u>925</u>	<u>172902</u>	<u>4/19</u>	<u>10.0</u>	<u>10.0</u>	<u>-207.8</u>	<u>(P) F</u>	<u>(L) F</u>
ICV	<u>"</u>	<u>1/25/18</u>	<u>930</u>	<u>176044</u>	<u>8/19</u>	<u>4.0</u>	<u>3.9</u>	<u>137.8</u>	<u>(P) F</u>	<u>(L) F</u>
CCV	<u>John Barrington</u>	<u>1/30/18</u>	<u>1310</u>	<u>172902</u>	<u>4/19</u>	<u>10.0</u>	<u>10.1</u>	<u>-211.7</u>	<u>(P) F</u>	<u>(L) F</u>
CCV									<u>P / F</u>	<u>L / F</u>

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 175.8, slope from 4 to 7 163 (both must be between 165 and 180 mV)

Is meter equipped with depth sensor? Yes / No If yes, list date of last quarterly depth verification

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value meters	Pass / Fail	Lab / Field
Pressure mode in air						<u>P / F</u>	<u>L / F</u>

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there is qualified data on this page.

Meter ID: 5

RQ: 2018-01-29-167

Project: CONT. MON. SMP

#FM

Notes: (1) Always wait for meter to stabilize before recording any readings.

(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).

(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Circle/Fill In: Lab Calibration or Calibration/Verification on Site

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 12/28/17

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	S. COFOR	1/30/18	0724	24.7	8.3	8.3	100	65	83755	⓪/F	⓪/F
ICV	" "	"	0722	24.7	8.3	8.3	100	64		⓪/F	⓪/F
CCV	S. COFOR	1/30/18	1344	23.2	8.5	8.7	100	63		⓪/F	⓪/F
CCV										P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Optical: DO gain range 0.85 to 1.15; DO charge N/A.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	S. COFOR	1/30/18	0725	170828	09/18	100	100	⓪/F	⓪/F
ICV	" "	"	0726	171211	12/18	100	985	⓪/F	⓪/F
CCV	S. COFOR	1/30/18	1344	270659	5/19	50	49615	⓪/F	⓪/F
CCV								P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	S. COFOR	1/30/18	0730	176157	8/19	7.0	7.0	-50	⓪/F	⓪/F
Calibr.	" "	"	0731	176044	8/19	4.0	4.0	121	⓪/F	⓪/F
Calibr.	" "	"	0733	172962	4/19	10.0	10.0	-224	⓪/F	⓪/F
ICV	" "	"	0735	172962	4/19	10.0	10.0	-223	⓪/F	⓪/F
CCV	S. COFOR	1/30/18	1346	176157	8/19	7.0	6.9	-47	⓪/F	⓪/F
CCV									P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 174, slope from 4 to 7 171 (both must be between 165 and 180 mV)

Is meter equipped with depth sensor? Yes ☒ No ☐ If yes, list date of last quarterly depth verification

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value meters	Pass / Fail	Lab / Field
Pressure mode in air						P/F	L/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS: