Balance Daily Use Record Mettler MS 403TS/00 Toxicity Lab
(Weights Bracketing Measured Weight)

STATE of FLORIDA
DEPARTMENT of ENVIRONMENTAL PROTECTION
BIOLOGY SECTION

Check Weight IDs
BIO-R-WEIGHTS-02 = R02
BIO-R-WEIGHTS-04 = R04
BIO-S-WEIGHTS-01 = S01

Mettler MS 403TS/00 serial # B815673069

Date	Material Weight or Weight Range	Lower Weight Value	Lower Weight Measured Value	Upper Weight Value	Upper Weight Measured Value	Check Weight ID	Weighed by: (Initials)	Comments

NOTE: Record g for grams; mg for milligrams
One lower value and one upper value check-weight can be used to bracket an event consisting of numerous measurements.
Additional check-weight measurements are required (per day) if more than two hours elapse between measurements.

Weight Tolerance - ≤ 1g = ± 2mg : > 1g ≤ 100g = ± 5mg : > 100g ≤ 400g = 14 mg

Comments: