Termination of Algal Assay Tests

1. SCOPE AND APPLICATION

1.1. This method is based on EPA 600/9-78-018, *The* Selenastrum capricornutum *Printz Algal Assay Bottle Test* and EPA 1974, *Marine Algal Assay Bottle Test*. It is used to determine growth of algal test species in freshwater and marine aqueous samples. Tests covered by this SOP are AA-AGP, AA-AGP-DCL, AA-AGP-F, AA-AGP-PH, AA-AGP-SA, AA-AGP-SPK, AA-AGP-SW, AA-LIM-NUT and AA-LN-NT-F.

2. SUMMARY OF METHOD

2.1. On various days during the 14 day test period, aliquots of inoculated samples are collected, diluted and counted with an electronic particle counter. Counts are uploaded into a data reduction program. Test endpoint is growth. Results are entered into the laboratory data management system.

2.2. Interferences include, but are not limited to, toxic substances introduced by contaminants in rinse water or on hands, glassware, sample hardware, and testing equipment. Dissolved oxygen <4.0 mg/L; pH <6.0 SU or >10.0 SU; salinity <1 part per thousand for freshwater tests; salinity <7 or >35 parts per thousand for marine tests; chlorine, debris, indigenous organisms, etc.

3. EQUIPMENT AND SUPPLIES

3.1. Fisher brand Dilu-Vials®, 30 mL

3.2. Coulter® Counter Model Z2 (S/N AB114302 and/or AF47542) - calibrated and used per SOP TA-06.19

3.3. Fisherbrand Bottletop Dispenser (Volume range 1-10 mL) - calibrated per SOP MB-03.3

3.4. 1 liter amber glass jar

3.5. Test-tube racks (to hold the 30 mL vials), 40 slots, as needed

3.6. Finnpipette II Pipettors (200-1000 µL adjustable volume and 1000 µL fixed volume) - calibrated per SOP MB-03.3

3.7. Pipettor tips, to fit Finnpipette II Pipettors

3.8. Black ink pen

4. REAGENTS AND STANDARDS

4.1. Deionized (DI) Water

4.2. Isoton®

5. SAMPLE COLLECTION, PRESERVATION, AND HANDLING

5.1. See SOP FS 1000 for sample collection and preservation.

The OFFICIAL COPY is the on-line version. All other copies are considered unofficial and uncontrolled.

6. SAMPLE PREPARATION

6.1. Refer to SOPs TA-08.02 and TA-08.03.

7. SAMPLE ANALYSIS

7.1. Fill a 1 liter amber glass jar with Isoton® and attach the Fisherbrand Bottletop Dispenser to the top. Perform three background counts on a sample of Isoton® dispensed from this jar. Record the counts and the mean cell volumes (MCV) on the *Background Sheet*.

7.1.1. Refer to SOP TA-06.19 to set up the Coulter® Counter and the LIMS LN AGP Manager.

7.2. For each sample flask that is to be counted, place one alpha-numerically numbered vial into the vial-holder rack. Use as many vials and racks as needed.

7.2.1. For the limiting-nutrient test, use your "shaker map" (SOP TA-08.03) to locate each flask to be drawn up.

7.3. Use the 1000 µL Finnpipette II pipettor to place 1 mL of sample from each test flask into separate 30 mL vials. Use a clean tip for each set of three flasks. Record the flask and vial codes on the *AGP/LN Count Record Sheet*.

7.4. Place 9.0 mL of Isoton® from the jar described in step 7.1 of this SOP into each vial containing samples.

7.4.1. Add the Isoton® to the vials one rack at a time to minimize settling of the algal cells. After one rack of vials has been counted, the Isoton® can be added to the next rack.

7.4.2. Wait for any bubbles that appear to disperse before counting the algal cells. This will prevent them from being included in the cell counts.

7.5. Record the count start time and analyst initials on the *AGP/LN Count Record Sheet*.

7.6. Perform three counts on the contents of each vial, as per SOP TA-06.19.

7.7. When finished with all of the counts, perform shut down procedures on the Coulter® Counter equipment, as per SOP TA-06.19.

7.8. After use, rinse the vials thoroughly with DI water. Shake to remove excess water and invert on dry paper towels to dry. Discard the vials after three to six months of usage or whenever they appear dirty or cracked, or if any contamination is noted.

7.9. On count day 14, after counting has been completed, remove the appropriate flasks from the shaker racks, empty the contents down the sink, and rinse the flasks with tap water. Place the flasks on the dirty glassware cart(s).

7.10. Analyze the raw data per SOP TA-08.07. Data is then corrected, verified, uploaded, verified again, and authorized in LIMS.

The OFFICIAL COPY is the on-line version. All other copies are considered unofficial and uncontrolled.

8. DATA ARCHIVAL

8.1. Use a black, waterproof pen for all bench sheets and logbooks.

8.2. Record counts in the appropriate bench sheet: *Background Counts Sheet* or *AGP/LN Count Record Sheet.*

8.3. To correct raw data entries, place a single line through the incorrect or illegible entry. Write the corrected entry near the error and initial the correction. Do not overwrite entries.

8.3.1. The corrected entry is written as near its appropriate space as possible or else in the comment box on the appropriate bench sheet.

8.3.2. If the correction requires an explanation, write a comment and reference the error. Initial and date the explanation.

8.3.3. Use sequential alphabetical annotation to identify errors and corresponding comments (use A for first error and use A for corresponding explanatory comment; use B for second error, etc.).

8.4. Enter the reduced raw data into the FDEP Laboratory Information Management System (LIMS).

9. QUALITY CONTROL

9.1. If a problem occurs during the sample processing or testing, the following steps are performed:

9.1.1. Lab supervisor is notified.

9.1.2. Client is notified (client makes decision for additional testing).

9.1.3. Section administrator is notified.

9.1.4. A non-conformance report (NCR) is submitted describing the problem and the corrective action.

9.1.5. To create an NCR: open **LIMS** and go to **Laboratory/Administrative/Non-Conformance Report**

9.2. Refer to http://www.dep.state.fl.us/water/sas/sop/index.htm for sample collection and preservation, and the Bureau Quality Manual for corrective action and handling out-of-control data.

10. SAFETY/HAZARDOUS WASTE MANAGEMENT

10.1. Proper laboratory protection, including lab hood or ventilation system, lab coat, gloves and safety glasses, is required when working with chemicals and unprocessed samples.

10.2. The shaker rack can snare loose clothing, especially lab coats, and can cause serious bodily injury. BE CAREFUL! The shaker racks are operated by a timer system that activates the shakers twice daily for a 60-minute period. The timer activates at 0800

The OFFICIAL COPY is the on-line version. All other copies are considered unofficial and uncontrolled.

and 1500. Always make sure the on/off switch is in the "off" position when placing flasks on and removing flasks from the shaker rack.

10.3. When replacing fluorescent lamps, make sure the shaker rack and the light switch are both turned off.

10.4. Refer to the FDEP Laboratory Quality Manual for sample disposal

10.5. Refer to the FDEP Laboratory Health and Safety Plan

10.6. Read MSDS information for reagent handling and disposal.

11. REFERENCES

11.1. EPA-600/9-78-018, *The* Selenastrum capricornutum *Printz Algal Assay Bottle Test*

11.2. EPA 1974, *Marine Algal Assay Bottle Test*

11.3. Inoculation of Algal Assay Tests, DEP SOP TA-06.19

11.4. Calibration and Use of the Coulter® Counter Model Z2, DEP SOP TA-06.19

11.5. Algal Growth Potential and Limiting Nutrient Data Verification and LIMS Upload, DEP SOP TA-08.07

11.6. FDEP Laboratory Quality Manual, http://www.dep.state.fl.us/labs/library/lab_sops.htm

11.7. FDEP Laboratory Health and Safety Plan, https://www.fldepnet.org/content/dear/bureau-laboratories-safety-documents

11.8. MSDS Information Sheets

APPENDIX OF CHANGES

08/28/2008	Added Scope and Application, Summary of the Method, Interferences, References and Appendix of Changes.
07/13/2009	Updated Quality Manual reference.
05/18/2010	Added "CAUTION Individuals allergic to latex should not use latex gloves. A suitable substitute should be found and used," and deleted "Vinyls gloves can be substituted for individuals allergic to latex" in Safety Section comments. Updated the reference to the Biology Quality Manual.
04/26/2011	Changed Biology Quality Manual reference to Bureau Quality Manual. Replaced "manipulate" with "reduce" and added "raw" to Step 9 of Counting an Algal Assay Sample.
02/06/2015	Formatted for 508 compliance.
03/15/2016	Updated references to FDEP Laboratory Quality Manual and FDEP Laboratory Health and Safety Plan

The OFFICIAL COPY is the on-line version. All other copies are considered unofficial and uncontrolled.