TIMER VERIFICATION
Culture Lab and Toxicity Incubators

1. SCOPE AND APPLICATION

1.1. These procedures are used to verify that prescribed light/dark cycles are properly controlled in the bioassay test chambers and the bioassay organism culture laboratory in accordance with requirements set forth in EPA 821-R-02-012, EPA 821-R-02-013 and EPA 821-R-02-014.

1.2. All incubator timers are set to 16 hours light and 8 hours dark (24-hour period) except those used for algal cultures, which are set for continuous lighting. Room lighting timer C is set to go off at 5:00 PM (to conserve energy).

2. SUMMARY OF THE METHOD

2.1. Current time as indicated by timers is compared to the correct time and adjusted (if necessary). The proper functioning of the on/off switches is tested by setting the timers near on/off times and allowing them to trip the switches. Supervisors are notified of any problems.

2.2. Timers are verified monthly, although the NELAC requirement is quarterly. Timers should also be checked and reset after power failure events. Verifications and aberrations are recorded in the *Light Timer Check* spreadsheet.

2.3. Daily observations of incubators (in use) and the culture laboratory are made and any aberrations from the scheduled cycle (lights off when scheduled to be on, etc.) should be reported and corrective action taken immediately.

3. SUPPLIES AND EQUIPMENT

3.1. Time piece (watch, clock, etc.)

3.2. *Light Timer Check* spreadsheet

3.3. Intermatic Model T173, 24-hour dial timer (Lab ID Timer C)

3.4. Dayton Model 4A214, 24-hour timer (Lab ID Timer A)

3.5. Dayton Model T1472BR, 24-hour dial timer (Lab Timer B)

3.6. Intermatic Model ET70815CR, Programmable Timer (Lab ID AGP Timer)

3.7. Toxicity Incubators #1 - #5

Note: Toxicity incubators are monitored per SOP TA-06.01

4. PROCEDURE

4.1. Mechanical Timers A, B, and C

4.1.1. Mechanical timers A and B are located on the west wall of B246; mechanical timer C is located in the southeast corner of B246.

The OFFICIAL COPY is the on-line version. All other copies are considered unofficial and uncontrolled.

4.1.2. Ensure that the time piece is within ± 1 minute of the correct time. Time can be checked with another time piece, computer, radio, etc. The U.S. Naval Observatory Master Clock (http://tycho.usno.navy.mil/cgi-bin/timer.pl) is a good source for the correct time.

4.1.3. Set the 'on' tab to 6:00 AM and the 'off' tab to 10:00 PM on each mechanical timer.

4.1.4. Set the mechanical timer to the current time by grasping the edge of the circular dial, pull outwards (about 1/4 inch) and turn clockwise until the current time mark on the dial is aligned as near as possible with the timer hand (silver colored stationary pointer-resembling an hour hand - located at the center of the dial that points towards the bottom of the dial).

4.1.5. To verify that the timer turns on the lights, set the timer to approximate 5:45 AM (regardless of the current time) by manipulating the circular dial as described in 4.1.4. Flip the manual lever (below the dial) to the left to turn off the controlled lights. The controlled lights should come on within 15 minutes.

4.1.6. To verify that the timer turns off the controlled lights, repeat 4.1.5., except set the time for 9:45 PM and leave the manual lever in the on position.. The controlled lights should turn off within 15 minutes.

4.2. **Incubators #1, #2, #3 #4 and #5)**

4.2.1. The timer for Incubator #2 (Marine AGP culture) is set for continuously on.

4.2.2. Incubators #1, #3 #4 and #5 are set for 16 hours light, 8 hours dark. Compare the time displayed on the display with the correct time. If the displayed time is not within ± 5 minutes of the correct time, use the procedures found in the Precision Instruction Manual to set the display to the correct time.

4.2.3. Follow the procedures in the instruction manual to check that the on/off times are set for 'on' at 6:00 AM and 'off' at 10:00 PM.

4.2.4. Check the 'off' time first (verification is normally done during the work day when the lights would be on). Follow procedures in the instruction manual to set the time to 9:55-59 PM. When the display shows 10:00 PM, the interior door lights should go off. Check by opening the door.

4.2.5. Check the 'on' time. Follow procedures in the manual to set the time to 5:55-59 AM. When the display shows 6:00 AM, the interior door lights should come on. Check by opening the door.

4.2.6. Record the date/time of verification on the *Light Timer Check* spreadsheet.

4.2.7. When a *Light Timer Check* spreadsheet is completed for the month, save the sheet as a PDF into \\fldep1\DEAR\\Labs\Lab5\Biology\Bench Common\ToxCommon\Instrument Certificates of Calibration\Light Intensity Meters\light timer verification.

The OFFICIAL COPY is the on-line version. All other copies are considered unofficial and uncontrolled.

4.3. Intermatic Model ET70815CR, Programmable Timer (Lab ID AGP Timer)

4.3.1. The Intermatic programmable timer is located on the east wall of B246 at the end of the AGP shaker platform. This timer controls the AGP shakers. It is set to operate for 1 hour twice daily during normal work hours (8:00 – 9:00 AM and 3:00 – 4:00 PM). Visually verify the operation of the shakers (should be 'on') during these periods. In case of power failure or if the displayed time is not within ± 5 minutes of the correct time, notify the lab supervisor. Refer to the Intermatic instruction manual for procedures to set the time of day and 'on' or 'off' periods.

5. DATA ARCHIVAL

5.1. Record timer checks in the the Excel spreadsheet *Light Timer Check*. When all monthly timer checks are completed and recorded, save the spreadsheet as a .pdf file named for the month being checked.

6. QUALITY CONTROL

6.1. Daily observations of incubators (in use) and the culture laboratory are made.

6.2. Timers are verified monthly by comparing current timer time to another time source to ensure it is within the acceptable range.

6.3. In case of power failure or if the displayed time is not within the acceptable range, the lab supervisor is notified and corrective actions are taken.

7. SAFETY/HAZARDOUS WASTE MANAGEMENT

7.1. Refer to the Laboratory Health and Safety Plan at https://www.fldepnet.org/content/dear/bureau-laboratories-safety-documents (Intranet)

7.2. Read MSDS information for reagent handling and disposal.

7.3. Use proper safety equipment including the lab hood or chemical ventilation system, a lab coat, safety glasses, a respirator and latex gloves (or a suitable substitute if allergic to latex) when working with chemicals and unprocessed samples.

8. REFERENCES

8.1. Operating Instructions for Intermatic Model T173, 24-hour dial timer

8.2. Operating Instructions for Dayton Model 4A214, 24-hour dial timer

8.3. Operating Instructions for Dayton Model T1472BR, 24-hour dial timer

8.4. Instruction Manual for Thermo-Precision Scientific Incubator Model 818

8.5. DEP SOP TA-06.01, Temperature Monitoring of Walk-in Refrigerator, Freezer and Toxicity Laboratoary Incubators

8.6. Instruction Manual for Intermatic Model ET70815CR Programmable Timer

The OFFICIAL COPY is the on-line version. All other copies are considered unofficial and uncontrolled.

8.7. EPA 821-R-02-012 *Methods for Measuring the Acute Toxicity of Effluents and Receiving Waters to Freshwater and Marine Organisms 5th ed.* October 2002.

8.8. EPA 821-R-02-013 *Short-term Methods for Estimating the Chronic Toxicity of Effluents and Receiving Waters to Freshwater Organisms 4th ed.* October 2002.

8.9. EPA 821-R-02-014 *Short-term Methods for Estimating the Chronic Toxicity of Effluents and Receiving Waters to Marine Organisms 3rd ed.* October 2002.

APPENDIX OF CHANGES:

07/25/2008	Added Scope and Application, Summary of the Method, References and Appendix of Changes
03/13/2013	Deleted Incubator #5. Added instructions for Intermatic Model ET70815CR, Programmable Timer (Lab ID AGP Timer)
03/13/2013	Reformatted the SOP.
03/03/2014	Updated the format of the SOP.
02/04/2015	Added incubator # 5, changed server information, and formatted for 508 compliance.
10/15/2015	Updated links. Minor corrections.
10/10/2016	Clarified timer light/dark cycle.

The OFFICIAL COPY is the on-line version. All other copies are considered unofficial and uncontrolled.