STANDARD PREPARATION PROCEDURE

1. SCOPE AND APPLICATION

This general procedure is followed when preparing standards used in the General Chemistry and Nutrients workgroups of the DEP Laboratory. It describes labeling and rotation of purchased standards, standard preparation procedures, documentation of prepared standards, labeling of standard containers, rotation of prepared standards, and disposal of used and expired standards. Any specific instructions about standard preparation that appear in the relevant SOP for the method shall take precedence over these instructions.

2. LABELING OF PURCHASED STANDARDS.

2.1. All shipping cartons containing standards for the General Chemistry and Nutrients workgroups should be labeled with the date of receipt by the receiving/purchasing staff. This information is needed for login to the LIMS Standard Prep Tracker after delivery of the standard to the analyst.

2.2. Label all individual standard containers with the following information:

- Unique ID number
- Expiration date (see below for exceptions)

All pertinent information about a standard is documented by logging it into the Standard Preparation Tracker in the Laboratory Information Management System (LIMS). A label is also printed using this module. If there is no expiration date supplied by the manufacturer for a purchased standard or no holding time stated in the method or SOP, enter “1/1/9999” in the Standard Prep Tracker’s entry field for the expiration date. In this case the module will print “None” on the label. If an expiration date is provided by the manufacturer it must be recorded on the container.

3. ROTATION OF PURCHASED STANDARDS.

Use the oldest dated standard first and replace with a newer dated stock standard as necessary. Rotate all purchased standards within the specified shelf life.

4. STANDARD PREPARATION INSTRUCTIONS.

Prepare standards as detailed in the relevant method SOP. The volumes prepared for the standards may be altered by the analyst as necessary. The final concentrations of the constituents of a standard must not be changed.

All standard preparations are documented in the reagent and standard logsheet (NU-RGTSTD-1.0). The logsheets are located at \\FLDEP1\DEAR\labs\lab3\Chem_data\Inorganic\Nutrients\REAGENT_PREP\ for each analysis. Submit the updated logsheet as part of the analysis report folder.

The OFFICIAL COPY is the on-line version. All other copies are considered unofficial and uncontrolled.

5. DOCUMENTING STANDARD PREPARATIONS.

Standards are given a unique ID by logging them in to the Standard Preparation Tracker in the Laboratory Information Management System (LIMS). Labels can be printed from this module. Certified standards can also be logged in the Standard Preparation Tracker, and any certification documentation supplied with such standards must be added to the Standards Logbook stored in room C307. Verification of stock and working standards is also documented in the LIMS Standard Preparation Tracker (see 8.3)

6. TRANSFERRING PREPARED STANDARDS.

Discard old standard from container. Wash standard container and lid six times with deionized water and allow to dry. Rinse standard container with a small volume of prepared standard and transfer remaining prepared standard to container.

7. LABELING PREPARED STANDARD CONTAINERS.

Label all prepared standard containers with the following information each time a standard is made:

- Unique ID number
- Identity
- Expiration date

All pertinent information about a prepared standard is achieved by logging it into the Standard Preparation Tracker in the LIMS and printing a label using this module.

8. ROTATION AND VERIFICATION OF PREPARED STANDARDS.

8.1. The holding times of all prepared standards are documented in the applicable SOP.

8.2. Use the oldest dated working standard first and replace with a newer dated working standard as necessary. Rotate all working standards within the specified holding times.

8.3 Make sure to prepare stock and working standards well before the ones in use expire. The newly prepared standards should be verified against the current ones and a standard reference material (SRM). For individual acceptance limits check the appropriate analysis SOP. Verification should be documented in the LIMS standard preparation tracker. Re-certification of standards needs approval from the supervisor.

8.3. Expired standards should be discarded as soon as possible after the date of expiration. Any expired standards used for non-quantification purposes should be labeled as such.

The OFFICIAL COPY is the on-line version. All other copies are considered unofficial and uncontrolled.

8.4. If standards are prepared for method development purposes only they must be labeled “For Research Only” or equivalent.

9. DISPOSAL OF USED AND EXPIRED STANDARDS.

9.1. Refer to the MSDS for information about chemicals being used.

9.2. Review the Laboratory Health and Safety Plan (accessible on the intranet at https://www.fldepnet.org/content/dear/bureau-laboratories-safety-documents).

9.3. Refer to the individual chemistry SOPs for specific handling of standards. Some standards used may need to be treated as hazardous or treated prior to disposal.

10. REFERENCES

10.1. Laboratory Health and Safety Plan.

10.2. Quality Manual, Bureau of Laboratories, FDEP.

Appendix of Significant Changes

08/24/2010 Instructions for the verification of standards are added to section 5 and 8.

The OFFICIAL COPY is the on-line version. All other copies are considered unofficial and uncontrolled.