AGP Light Intensity Measurement Log (SOP TA-06.26)

Name: _____
Date: _____
Meter # _____

Initial Readings					
Shaker	Left	Right	Front	Back	Bulbs Replaced
A □					
B □					
C □					
D □					
E □					
F □					

Incubator # 7 Precision Readings (Serial # WB80269161)

Readings After Lightbulb Changes				
Shaker	Left	Right	Front	Back
A				
B				
C				
D				
E				
F				

Comments:

Notes:
Meter readings should be taken at sample level and at appropriate edge of rack. Acceptable range is 396-484 FC. When corrected for flourescent lighting, this is 360-440 FC. Bulbs are numbered from front to back.