Bioassay Sample Log-In Procedure

1. SCOPE AND APPLICATION

    1.1. This method is used to document the receipt and the physical-chemical characteristics of bioassay samples received in the bioassay laboratory.

2. SUMMARY OF METHOD

    2.1. The transfer of bioassay samples from the Laboratory Receiving room to the bioassay lab is documented on the chain of custody form (sample submittal form). Bioassay sample collection and receipt information is logged into the *Bioassay Sample Log-in* book. Physical-chemical parameters are measured on an aliquot of the bioassay sample. The results of these measurements are entered into the *Bioassay Sample Log-in* book and LIMS logbook. Labels are applied to the sample containers. The LIMS Custody module is used to track sample custody.

3. EQUIPMENT AND SUPPLIES

    3.1. YSI MultiLab 4010-1 (SOP TA-06.44)

    3.2. Fisherbrand Traceable Bench Conductivity Meter (SOP TA-06.36)

    3.3. Beaker or specimen cups

    3.4. Digital thermometer (SOP TA-06.13)

    3.5. Fisherbrand Accumet AB315 pH/mV Meter (SOP TA-06.43)

    3.6. Hach Pocket Colorimeter II Test Kit for Total Residual Chlorine (SOP TA-06.22)

    3.7. Hach Ammonia Nitrogen Pocket Colorimeter II Test Kit (SOP TA-06.35)

    3.8. Titration Kits and Reagents for Alkalinity and Hardness (SOP TA-06.08)

    3.9. Bucket or plastic Carolina dish for water bath

    3.10. Black, waterproof pen

    3.11. Computer with access to the Laboratory Information Management System (LIMS)

    3.12. *Bioassay Sample Log-in* logbook

    3.13. *Bioassay bench sheets* logbooks

4. REAGENTS AND STANDARDS

    4.1. Aliquot of bioassay sample

5. METHOD

    5.1. Sample Receipt

The OFFICIAL COPY is the on-line version. All other copies are considered unofficial and uncontrolled.

5.1.1. Sign the chain-of-custody form (sample submittal form) when picking up a sample from Laboratory Receiving Room. Scan the employee badge of the Receiving Room staff that transfers the sample.

5.1.2. Record the following in *Bioassay Sample Log-in* book:

a. Facility name

b. LIMS Job number and LIMS Sample ID (both are usually obtained the day after the sample is received in the bioassay lab). Use the RQ on the sample bottle to search LIMS Browser for the Job and Sample IDs. See 5.2.1 for additional information.

5.1.3. Record the time and date the sample was collected.

5.1.4. Transfer an aliquot of sample going to be used the same-day into an appropriate flask and place in a warm-water bath (sink or bucket). Allow the sample to warm to the appropriate test temperature (usually 25 ±1 °C). Place the rest of the sample set into the walk-in refrigerator until needed.

5.1.5. Record the date and time sample was received in the Toxicity section.

5.1.6. Record the name of the Biology staff member who picked up the sample from the Receiving Room.

5.1.7. After warming, pour ~150 mL of the sample into a specimen cup and measure the following: temperature, DO, pH, conductivity, salinity, chlorine, ammonia, alkalinity, and hardness. If chlorine is measured to be above 0.1 mg/L, dechlorinate the sample according to SOP TA-09.18.

5.1.7.1. Record temperature, DO, pH, conductivity and salinity in the *Bioassay Sample Log-in* book and LIMS (per SOP TA-06.42). Record chlorine, ammonia, alkalinity, and hardness in their respective logbooks and LIMS logbook.

5.1.8. Conversely, a cup of sample can be poured from the bioassay jug prior to being warmed and then warmed separately to speed up the login process.

5.2. Sample Label Receipt

5.2.1. Job Folders are no longer prepared by Receiving Staff. Biology staff can get sample labels from the Biology Box in the Receiving Room. Request that the receiving room staff print a label (containing identical information) for each bioassay sample container from the same facility. These can also be found by using LIMS browser.

5.2.2. Record Sample ID and Job numbers from the Sample Label in the *Bioassay Sample Log-in* book and on the appropriate bioassay bench sheets. Place one label on each of the sample containers.

5.2.3. Review field sheet. Verify the date and time of sample collection.

The OFFICIAL COPY is the on-line version. All other copies are considered unofficial and uncontrolled.

5.2.3.1. Field sheets are now scanned by Receiving Staff. To view the field sheet, go to BizhubScans. The Receiving paperwork (field sheets) are saved by Event ID. The Event ID can be obtained in LIMS Browser using the sample ID as the search parameter.

5.2.4. Verify that all samples are listed on field sheets.

5.2.5. Verify that all samples have a LIMS sample ID assigned.

5.2.5.1. Write the LIMS sample ID on the bench sheet.

5.2.6. Verify that all samples received have the correct analysis or analyses requested and are logged into LIMS.

5.2.6.1. Notify Receiving Room Staff as soon as practical to make corrections if an error or omission is found.

5.2.7. Verify that all samples received had the correct sample and station/field IDs assigned to them during log-in as listed on the field sheet.

5.2.7.1. Notify Receiving Room Staff to make corrections.

5.2.8. Verify that all samples have a sample date and time listed.

5.2.8.1. Make sure the sample date/time are the end date/time it was collected if it was a composite sample. Record sample date/time in the *Bioassay Sample Log-in* book.

5.3. Sample Disposal

5.3.1. When a sample is depleted or a test is completed, triple rinse the sample containers with tap water, and throw away in the recycling bin.

6. DATA ARCHIVAL

6.1. Use a black, waterproof pen for all bench sheets and logbooks.

6.2. Record information in *Bioassay Sample Log-in* book, LIMS logbook and appropriate *Bioassay bench sheets* logbooks.

6.3. FDEP Laboratory Information Management System (LIMS)

6.4. To correct raw data entries, place a single line through the incorrect or illegible entry, write the corrected entry near the error, date and initial the correction. Do not overwrite entries.

6.4.1. The corrected entry is written as near as possible to the error or in the comment box on the appropriate bench sheet.

6.4.2. If the correction requires an explanation, write a comment and reference the error with a letter. Initial and date the explanation. Do not attempt to put all information (correction and comments) into a small space.

The OFFICIAL COPY is the on-line version. All other copies are considered unofficial and uncontrolled.

6.4.3. Use sequential alphabetical annotation to identify errors and corresponding comments (use A for first error and use A for corresponding explanatory comment; use B for second error, etc.).

7. QUALITY CONTROL

7.1. Refer to DEP Laboratory Quality Manual for corrective action and handling out-of-control data.

7.2. Due to the short-hold time of the bioassay samples, organism age constraints, sample manipulation and preparation, among other issues, initial sample custody by Biology staff begins when the samples are transferred to Biology staff as noted by Biology staff signing the Sample Submittal forms (aka field sheets or chain of custody forms) and noting the number of bioassay sample jugs taken. Samples are taken directly to the bioassay lab for preparation. Excess sample and/or sample containers are placed into the walk-in refrigerator in Room B227.

8. SAFETY/HAZARDOUS WASTE MANAGEMENT

8.1. Proper laboratory protection, including lab hood or ventilation system, lab coat, gloves and safety glasses, is required when working with chemicals and unprocessed samples.

8.2. Refer to Laboratory Quality Manual for sample disposal.

8.3. Refer to the Laboratory Health and Safety Manual and Hazardous Waste Management.

9. REFERENCES

9.1. DEP SOP TA-06.08, *Measuring Total Alkalinity and Hardness*

9.2. DEP SOP TA-06.13, *Thermometer Accuracy Check and Use*

9.3. DEP SOP TA-06.22, *Measuring Total Residual Chlorine using the Hach Pocket Colorimeter and Colorimeter II Test Kits*

9.4. DEP SOP TA-06.35, *Measuring Total Ammonia using the Hach Ammonia Nitrogen Pocket Colorimeter II Test Kit*

9.5. DEP SOP TA-06.36, *Calibration and Use of the Fisherbrand Traceable™ Bench Conductivity Meter for Conductivity and Salinity*

9.6. DEP SOP TA-06.44, YSI MultiLab 4010-1

9.7. DEP SOP TA-06.43, *Calibration and Use of Fisherbrand accumet AB315 pH/mV Meter*

9.8. DEP SOP TA-09.18, *Dechlorination Procedure using Sodium Thiosulfate*

9.9. DEP SOP TA-06.42, *LIMS Entry for Laboratory Parameters*

The OFFICIAL COPY is the on-line version. All other copies are considered unofficial and uncontrolled.

9.10. Laboratory Quality Manual for sample disposal https://floridadep.gov/dear/florida-dep-laboratory/content/dep-laboratory-quality-assurance-manual-and-sops

Laboratory Health and Safety Manual and Hazardous Waste Management https://www.floridadep.net/resources/health-safety-manual

APPENDIX OF CHANGES

01/02/2008	Added Scope and Application; Summary of Method; and Appendix of Changes.
12/03/2008	Added References.
08/12/2013	Reformatted SOP.
01/30/2015	Formatted for SOP compliance.
07/09/2015	Updated links.
07/11/2016	Updated links.
08/28/2018	Updated links and changed font to Calibri.
09/13/2018	Added new SOPs to references and new pH and conductivity meters in equipment list. Renewed Bizhubscans link. (M. Faircloth 09/13/18)
07/31/2019	Updated meters and SOPs. Added section on checking sample collection end date and time on field sheet. Added section on measuring chlorine, ammonia, alkalinity and hardness. Renewed Bizhubscans link.
09/01/2020	Updated equipment section. Added LIMS mentions. Updated data archival section with universal language.
07/07/2021	Checked for grammar issues. Updated names for SOPs in references section.
08/11/2022	Updated pH meter. Updated links.
08/14/2023	Updated DO meter. Checked readability.
2/2/2024	Removed unnecessary second temperature check.

The OFFICIAL COPY is the on-line version. All other copies are considered unofficial and uncontrolled.